

> Denotes revised material in this JPM

JOB PERFORMANCE MEASURE

TASK CODE: CFO-130

TASK: Operate an Alarming Dosimeter (EPD) and REMS Computer

NAME: _____ **BADGE:** _____

>**REFERENCES:** Only references for knowledge items are listed here. The trainee is expected to identify the correct references for practical items.

1. Electronic Personal Dosimeter Handbook (ESL)
2. User Guide for REMS Manager Workstation

TERMINAL OBJECTIVE:

Given that an alarming dosimeter is required for use, operate the alarming dosimeter per WP 12-HP1400.

CONSEQUENCES OF INADEQUATE PERFORMANCE:

Excessive personnel exposure

HAZARDS (PERSONNEL/EQUIPMENT STATUS):

Personnel overexposure

PRE-REQUISITE TRAINING/ TASK COMPLETION:

1. CL 1.00 Series
- >2. CL 2.04, Dosimetry

TOOLS/EQUIPMENT (MATERIALS REQUIRED):

- | | |
|-----------------------------|-----------------------------|
| 1. Alarming dosimeter (EPD) | 2. REMS Manager Workstation |
| >3. EPD batteries | |

Instructions to Trainee: You shall acquire the necessary references and equipment, and complete all required documentation. Knowledge requirements shall be completed with 80% or greater accuracy. Critical step performance shall be completed with 100% accuracy.

> Denotes revised material in this JPM

Instructions to JPM Evaluator: The trainee is to perform the terminal objective, without assistance, on the job site. Provide clarification of requirements if requested by the trainee. You are encouraged to ask relevant questions to verify trainee understanding. If the trainee fails this JPM, clearly document the reason for failure and forward to the trainee's manager. Successful completion of this JPM shall be recorded on the trainee's qualification card.

>On performance items, if there is more than one means available to accomplish the step, circle the method used. Refer to the RCT-01 Qualification Standard for preference of use.

KNOWLEDGE REQUIREMENTS:

Reference	Knowledge Requirement	Pass/Fail
1	Describe the basic operation of an alarming dosimeter.	
1	Discuss the capabilities of the alarming dosimeter.	
>2	Discuss operations that can be performed on the REMS Manager Workstation	

PERFORMANCE REQUIREMENTS:

>Method	Performance Requirement	Pass/Fail
P	Perform a battery change out. #	
P	Set the dose rate alarm on the REMS Manager Workstation. #	
P	Set the dose alarm on the REMS Manager Workstation #	
>P	Enter RWP information into the REMS Manager Workstation. #	
>P	Assign worker(s) to an RWP in the REMS Manager Workstation. #	
>P	Retrieve reports from the REMS Manager Workstation.	

indicates a critical step

> Denotes revised material in this JPM

FINAL EVALUATION:

PASS

FAIL

COMMENTS:

EVALUATOR SIGNATURE:

_____ **DATE:** _____

TRAINEE SIGNATURE:

_____ **DATE:** _____

MANAGER SIGNATURE:

_____ **DATE:** _____